



## **FDAS SUPPLEMENTARY CRITERIA FOR CALIBRATION LABORATORIES**

### **1.0 Introduction**

**1.1. Scope:** These criteria are for Calibration Laboratories.

#### **1.2. Normative and other Reference**

**1.2.1.** ISO/IEC 17025:2017, General Requirements for the Competence of Testing and Calibration Laboratories.

**1.2.2** ISO/IEC 17000:2000 Conformity Assessment – Vocabulary and General Principles

**1.2.3.** ILAC-P9:06/2014 ILAC Policy for Participation in Proficiency Testing Activities

**1.2.4.** ILAC-P10:07/2020 ILAC Policy on the Traceability of Measurement Results

**1.2.5.** ILAC-P14:09/2020 ILAC Policy for Uncertainty in Calibration

### **2.0 Definitions**

Applicable definitions of ISO/IEC Standard 17000 series apply.

### **3.0 Proficiency Testing Activity**

The minimum amount of appropriate proficiency testing required per laboratory is given below:

- One PT/ILC activity from each applied field prior to gaining accreditation.
- One PT/ILC activity relating to each group of laboratory's scope of accreditation at least every four years.

### **4.0 Links to Additional References**

4.1 Asia Pacific Accreditation Cooperation-[www.apac-accreditation.org](http://www.apac-accreditation.org)

4.2 International Laboratory Accreditation Cooperation – [www.ilac.org](http://www.ilac.org)

4.3 International Organization for Standardization – [www.iso.org](http://www.iso.org)

4.4 International Electrotechnical Commission – [www.iec.ch](http://www.iec.ch)