



Federation for Development of Accreditation Services

CERTIFICATE OF ACCREDITATION (AS PER ISO/IEC 17025:2017)

This is to attest that

M/s Meya Labs Pvt. Ltd.

No. 49, Ground Floor, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur
Chennai (TN)-600077

Calibration Laboratory


has demonstrated compliance with ISO/IEC Standard 17025:2017, General requirements for the competence of testing and calibration laboratories, and supplementary criteria for calibration laboratories.

Certificate Number: CL-104

Issue Date: 28/05/2022

Valid Until: 27/05/2024

The certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard and the relevant requirements of FDAS (scope of accreditation also displayed at www.fdasindia.org).


DEVI SARAN TEWARI
President



SCOPE OF ACCREDITATION

Laboratory Name M/s Meya Labs Pvt. Ltd.
No. 49, Ground Floor, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur
Chennai (TN)-600077

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CL-104 Page No.: 1/2

Validity 28/05/2022 to 27/05/2024 Last Amended on

MECHANICAL CALIBRATION

Laboratory based Calibration

S. No.	Parameter	Standard(S), BIS, ISO, ASTM, Regulators, etc.	Range	Uncertainty In Measurement (±)
1	Torque: Torque Generating Device (Rotary Torque tools, Pneumatic, Electrical & Hydraulic Tools)	IS 16906:2018	0 to 50 kNm	1.8%
2	Torque: Torque Measuring Device (Torque Sensor with or without indicator/Torque Calibrator)	BS-7882:2018	10 to 1000 Nm	1.2% for 10 Nm to 200Nm 0.4% for 200Nm to 1000Nm
3	Pressure: (Absolute Pressure Indicator, Barometer, Absolute pressure gauges, Absolute transmitter, Absolute Transducer)	DKD R-6-1	500 to 1000 mbar (Abs)	2.4 mbar


Dealing Officer



Federation for Development of Accreditation Services

SCOPE OF ACCREDITATION

Laboratory Name M/s Meya Labs Pvt. Ltd.
No. 49, Ground Floor, Moorthy Nagar, 3rd Street, Chettiar Agaram, Porur
Chennai (TN)-600077

Accreditation Standard ISO/IEC 17025:2017

Certificate Number CL-104 Page No.: 2/2

Validity 28/05/2022 to 27/05/2024 Last Amended on

OPTICAL & PHOTOMETRY CALIBRATION

Laboratory based Calibration

S. No.	Parameter	Standard(S), BIS, ISO, ASTM, Regulators, etc.	Range	Uncertainty In Measurement (±)
1	Luminous Intensity: Lux Meter, Light Meter, Illumines Meter	Inhouse Calibration Method	100 to 10000 Lux	4%


Dealing Officer