



FDAS SUPPLEMENTARY CRITERIA FOR CALIBRATION LABORATORIES

1.0 Introduction

1.1. Scope: These criteria are for Calibration Laboratories.

1.2. Normative and other Reference

1.2.1. ISO/IEC 17025:2017, General Requirements for the Competence of Testing and Calibration Laboratories.

1.2.2 ISO/IEC 17000:2000 Conformity Assessment – Vocabulary and General Principles

1.2.3. ILAC-P9:06/2014 ILAC Policy for Participation in Proficiency Testing Activities

1.2.4. ILAC-P10:07/2020 ILAC Policy on the Traceability of Measurement Results

1.2.5. ILAC-P14:09/2020 ILAC Policy for Uncertainty in Calibration

2.0 Definitions

Applicable definitions of ISO/IEC Standard 17000 series apply.

3.0 Policy on Proficiency Testing:

FDAS requires its laboratories to comply with FDAS 135, FDAS Policy on Proficiency Testing.

4.0 Policy on Metrological Traceability:

Laboratory to establish metrological traceability of measurement results by practicing FDAS 136, FDAS Policy on Metrological Traceability.

5.0 Policy on Shifting of CAB's Location:

In case laboratory wants to shift to new location after grant of accreditation, it should comply with FDAS 137, FDAS Policy on Shifting of the CABs location.

6.0 Links to Additional References

6.1 Asia Pacific Accreditation Cooperation-www.apac-accreditation.org

6.2 International Laboratory Accreditation Cooperation –www.ilac.org

6.3 International Organization for Standardization – www.iso.org

6.4 International Electrotechnical Commission – www.iec.ch