



FDAS Policy on Metrology Traceability

Policy: The Metrology Traceability policy of FDAS for its accredited CABs is to ensure traceability to SI Units and by abiding with the requirements of ILAC- P10: 07/2020 "ILAC Policy on Metrology Traceability of Measurement Results", by practicing the following:

1. FDAS accepts the traceability to measurements of the measuring devices that have been calibrated by NPL- New Delhi, which is India's National Metrology Laboratory, and maintains its protocol of equivalence with BIPM/OIML/RMOs (say APMP) etc., by abiding required norms.
2. FDAS also accepts the traceability to measurements of the measuring devices that have been calibrated by the laboratories having accreditation from National or from International Accreditation bodies that are operating their management system as ISO/IEC 17011: 2017 and are having MRA / full member status.
3. FDAS accepts the traceability to chemical measurements achieved through CRMs procured from of the accredited CRMs/BNDs*/RMs provider.
4. In those situations where a laboratory may not be in position to comply with above conditions but refers to with the compliance of ILAC P-10, FDAS would assess and evaluate that situation to arrive at its decision.

BNDs* - Bhartiya Nirdeshak Drabya.